



# CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 90-001108 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

By  
6/26/09

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 6/5/09 Cal 1 N/A AC Cal 2 N/A AC

Between the dates of - and -  
the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 6/5/09.  
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione


6/25/09

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 25<sup>th</sup>, day of June, 20 09.

A.M. Hernandez

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.



TUCSON POLICE DEPARTMENT  
CRIME LABORATORY

1 of 2x

INTOXILYZER SERVICE RECORD

by  
6-26-09

Date: 6/18/09

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-001108

Reason For Service: Periodically fails the Modern Test during initial  
Diagnostics

Date Out of Service: 6/5/09

Calibration Check completed prior to removal from service

Date: 5/28/09

SQAP completed prior to removal from service

Date: 5/28/09

Corrective Action:

☒ Sent to CMI, Inc.

Data downloaded prior to sending to CMI

2 of 2sc

TSY  
6-26-09



## Return Material Authorization Form

This form **MUST** be completed and enclosed with item(s) being returned.  
*Failure to complete and return this form may cause delays in crediting / repairing.*

1. Item(s) Returned: ☒ Instrument ☐ Supplies ☐ Other  
(Note: please ship items in their original shipping container(s) or a similar protective box.)

2. Instrument Model: Intox 8000 Serial Number 80-001108

3. Bill To Address: Finance / Accounts Payable  
P.O. Box 27450  
TUCSON, AZ 85726-7450

Ship To Address: (NO PO BOXES)  
Tucson Police Department  
Crime Lab  
270 S. Stone Ave  
TUCSON, AZ 85701-1917

Customer # 105075

Name Terry Gallegos

Phone: (520) 837-7677 Fax: (520) 791-4642

Email: Terry.Gallegos@tucsonaz.gov

4. Reason for Return: Note - If you are sending an item for repair, please give a detailed description of the problem. Please list any special instructions that you may have concerning this return. Note - Returns for credit may be subject to a 15% restocking charge.

Periodically fails the Modern Test during the initial diagnostics

5. Purchase Order Number (attach a copy of P.O. if applicable) 20669

6. I Authorize All Repairs: ☐ Yes ☐ No

7. I Authorize Repairs Up To: \$                     

8. I Require An Estimate Regardless of Cost ☒ Yes ☐ No

**Note: An evaluation fee will apply to estimates that are not repaired.**

9. Authorized By:

Terry Gallegos  
Name (Please Print)

Signature

Date

6-5-09

Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department